

## REQUEST FOR QUOTATION

RFQ Reference: **RFQ-MD058-25**

Date: **10 November 2025**

Subject of RFQ): **UVC recirculatory system**

International Organization for Migration kindly requests your quotation for the provision of the goods described in the RFQ submission form below.

When preparing your quotation, please be guided by the RFQ information below. It is your responsibility to ensure that your quotation is submitted on or before the deadline. Quotations received after the submission deadline, for whatever reason, will not be considered for evaluation.

### RFQ INFORMATION

<b>Deadline for the submission of quotation</b>	19 November 2025
<b>Method of submission</b>	Quotation must be submitted as follows: <input type="checkbox"/> E-tendering <input checked="" type="checkbox"/> Email
<b>Cost of preparation of quotation</b>	IOM shall not be responsible for any costs associated with a vendor’s preparation and submission of a quotation, regardless of the outcome or the manner of conducting the selection process.
<b>Contractual Terms</b>	Any Purchase Order that will be issued as a result of this RFQ shall be subject to the IOM standard terms for provision of goods/services/transportation/medical services available at <a href="https://www.iom.int/do-business-us-procurement">https://www.iom.int/do-business-us-procurement</a> or IOM standard contract templates.
<b>Documents to be submitted</b>	Bidders shall submit and sign the-bid submission form below.
<b>Quotation validity period</b>	The quotation shall remain valid for <input type="text"/> days from the deadline for the submission.
<b>Price</b>	Quotations shall preferably be submitted in MDL, and be be exclusive of VAT and other applicable indirect taxes. If a different currency is proposed, the evaluation will be conducted using the current IOM exchange rate, available at <a href="https://treasury.un.org/operationalrates/OperationalRates.php">https://treasury.un.org/operationalrates/OperationalRates.php</a> .
<b>Payment terms</b>	<input checked="" type="checkbox"/> 100% within 20 days after receipt of goods, works and/or services and submission of payment documentation.
<b>Partial quotations</b>	<input checked="" type="checkbox"/> Not permitted
<b>Right to vary requirement at time of award</b>	At the time of award of Contract or Purchase Order, IOM reserves the right to vary (increase or decrease) the quantity of services and/or goods, by up to a maximum 25% of the total offer, without any change in the unit price or other terms and conditions.
<b>Clarifications</b>	Contact person for correspondence, notifications and clarifications Contact person: Ina DABIJA E-mail address: idabija@iom.int
<b>Evaluation method</b>	<input checked="" type="checkbox"/> The contract will be awarded to the lowest price substantially compliant offer
<b>Right not to accept any quotation</b>	IOM is not bound to accept any quotations, nor award a contract or purchase order
<b>Expected date for contract/PO award.</b>	01.12.2025

Issued by:

Signature: \_\_\_\_\_

## Requirements

<b>Air disinfection system (UVC Recirculatory) high capacity- 100 pcs</b>	
<b>Description</b>	Device for disinfecting indoor air by destroying viruses, bacteria, and fungi in the presence of patients and medical staff.
<b>Parameter</b>	<b>Specification</b>
<b>Actual processing volume</b>	$\pm 30 \text{ m}^3/\text{h}$
<b>Operating principle</b>	Air flow recirculation and microorganism destruction
<b>Microorganism destruction location</b>	Inside the device
<b>Core technology</b>	UVC (Ultraviolet C) rays
<b>Type of bactericidal lamps supported</b>	Lamp type: T8 (double-ended);
	Connectors: G13 (two-pin);
	Lamp power: minimum 55W;
	Lamp length: approx. 900 mm
<b>Effective working life of UV lamps</b>	min 9000 h
<b>Recirculator equipped with bactericidal lamps</b>	yes
<b>Bactericidal lamps installed inside the recirculator upon delivery</b>	yes
<b>Use of the device in the presence of patients and medical staff</b>	No restrictions
<b>Presence of by-products during operation</b>	No ozone or other by-product emissions
<b>Device casing</b>	Stainless steel (minimum AISI 304)
<b>Maximum noise level</b>	$\leq 56 \text{ dB}$
<b>Number of air recirculation fans</b>	min 1
<b>Operating hours counter</b>	Mandatory
<b>Remote control</b>	Mandatory
<b>Timer module</b>	yes
<b>Protection against overvoltage or short circuit</b>	yes
<b>Equipped with glass, plastic windows or indicators for operation visibility</b>	yes
<b>Energy consumption</b>	$\leq 400 \text{ W/h}$
<b>Power supply voltage</b>	220-240VAC, 50Hz
<b>Mounting method</b>	Wall (vertical/horizontal) or ceiling
<b>Laboratory tests for disinfection efficiency</b>	yes
<b>Device labeling according to Agency for Medicines and Medical Devices of the Republic of Moldova (AMDM) recommendations</b>	yes

<b>User manual</b>	Yes, in the official state language
<b>Warranty</b>	min 24 months

Thank you and we look forward to receiving your quotation.